

GE Healthcare

# Valve INV-907

Instructions





## Important user information

---

All users must read this entire manual to fully understand the safe use of Valve INV-907.

### WARNING!

---



The WARNING! sign highlights instructions that must be followed to avoid personal injury. It is important not to proceed until all stated conditions are met and clearly understood.

### CAUTION!

---

The Caution! sign highlights instructions that must be followed to avoid damage to the product or other equipment. It is important not to proceed until all stated conditions are met and clearly understood.

### Note

---

The Note sign is used to indicate information important for trouble-free and optimal use of the product.

## CE Certifying

---

This product meets the requirements of applicable CE-directives. A copy of the corresponding Declaration of Conformity is available on request.

The CE symbol and corresponding declaration of conformity, is valid for the instrument when it is:

- used as a stand-alone unit, or
- connected to other CE-marked GE Healthcare instruments, or
- connected to other products recommended or described in this manual, and
- used in the same state as it was delivered from GE Healthcare except for alterations described in this manual.

## Recycling

---



This symbol indicates that the waste of electrical and electronic equipment must not be disposed as unsorted municipal waste and must be collected separately. Please contact an authorized representative of the manufacturer for information concerning the decommissioning of equipment.

### WARNING!

---

This is a Class A product. In a domestic environment this product may cause radio interference in which case the user may be required to make adequate measures.



# Contents

---

1	Introduction .....	7
2	Installation .....	9
3	Operation .....	11
4	Maintenance .....	15
5	Trouble-shooting .....	17
6	Reference information .....	19



# 1 Introduction

---

Valve INV-907 is a motorised rotary 7-port valve. The valve is used in ÄKTA™ design chromatography systems. It is powered from the ÄKTA design P-900 series system pump, and is controlled from UNICORN™ control system.

The valve has 3 positions used for:

- loading a sample loop
- injecting the sample onto the column
- washing the system pump.

The valve can also be used to switch between upflow and downflow in a column.

## Features:

- Completely swept flow path minimises eluent or sample “memory effect”.
- Flow rates up to 100 ml/min.
- All wetted parts are plastic PEEK.

## Safety



**WARNING!** When using hazardous chemicals, take all suitable protective measures, such as wearing protective glasses and gloves resistant to the chemicals used. Follow local regulations and instructions for safe operation and maintenance of the system.

## 1 Introduction



## 2 Installation


**CAUTION!** Before connecting Valve INV-907 ensure the power is switched OFF at the system pump or the complete system.

### Unpacking


Unpack the valve and check the items against the packing list. Inspect the items for obvious damage which may have occurred during transportation.

### Installing the valve

- 1 Connect the valve with two UniNet cables as a part of the UniNet 2 chain.
- 2 Set the ID-switch at the bottom to the required valve number 0-9. The number should correspond to that used in UNICORN.




Uninet 2 network



**Note:** All valves must have different numbers.

- 3 Mount the valve vertically, in the way it is shown in the user documentation of your ÄKTAdesign chromatography system.
- 4 Apply an identification number from the labels supplied. The number should be the same as that of the rotary switch at the bottom of the valve. Place the number so that it is easily readable from the front.
- 5 Connect the tubing.



## 2 Installation

6 When used as an injection valve connect the tubing as follows:

<b>Port</b>	<b>Connect to</b>
1	Column inlet
2	Sample loop or bottom tubing of Superloop
3	Injection port
4	Waste
5	Waste
6	Sample loop or top tubing of Superloop
7	Outlet of the pump

7 When used as an upflow/downflow valve, connect the valve as indicated in the user documentation of your ÄKTAdesign chromatography system.


## 3 Operation

The valve is controlled from UNICORN. When used as an injection valve use the instruction **InjectionValve** in **System Control:Manual:Flowpath**. The valve can be set in 3 positions:


- LOAD position 1 (home position)
- INJECT position 2
- WASTE position 3

If the valve has not been activated it is in position 1.


### LOAD, position 1



### INJECT, position 2




### WASTE, position 3




### **Manual sample loading**

The sample loop and Superloop can be loaded with a syringe. Connect Superloop with the bottom tubing connected to port 2 and the upper tubing to port 6.

For large volumes use a luer female/1/16" male union connected to port 3.




For smaller volumes use an injection needle together with the injection fill port as described below:



Check the fill port nozzle is screwed into the fill port.

**Note:** The thread is left-handed.

- 1 Loosely thread the fill port screw into valve port 3.
- 2 Insert the injection needle (0.7 mm o.d.) into the injection fill port.
- 3 Tighten the fill port until the nozzle has formed a seal around the needle tip. When the seal is adjusted correctly, it feels as if you are penetrating a septum at the end of the injection fill port. The seal should provide easy insertion and removal of the needle.
4. Mount the syringe holder on the fill port.



Two techniques can be used for filling the sample loop; partial or complete filling.

Type of filling	Volume to load
Partial filling	max 50% of the sample loop volume
Complete filling	2–5 times the sample loop volume

### Partial Filling

Partial filling is used when high recovery is required. Partial filling allows the injected volume to be changed without changing the loop and does not waste sample.

Partial filling is achieved as follows:

- 1 Set injection valve to position LOAD.
- 2 Load the syringe with a large volume of buffer (5× the loop volume).
- 3 Fill the sample loop carefully with buffer.
- 4 Set the injection valve to position INJECT before taking out the syringe.

**Note:** If the syringe is taken out when the injection valve is in position LOAD, self drainage will occur and air will enter the sample loop.

- 5 Load the syringe with the required volume of sample. No more than half (50%) a loop volume of sample should be loaded into the loop.
- 6 Insert the syringe in position 3 on the injection valve. Set the injection valve to position LOAD.

**Note:** Do not load the sample before the valve is in position LOAD.

- 7 Gently load the syringe contents into the loop.
- 8 Leave the syringe in position. The sample will be injected onto the column when the valve is switched to INJECT in the method.

### Complete filling

In this method a surplus of sample is used to ensure that the sample loop is filled completely, allowing reproducible sample volumes. In preparative applications the sample volume should be at least 2 times the volume of the sample loop. For analytical reproducibility a sample volume 5 times the volume of the sample loop should be used.

With complete filling the sample volume can only be changed by changing the loop size.

Complete filling is achieved as follows:

- 1 Set valve to position LOAD.
- 2 Load the syringe with sample (2-5 times the loop volume).
- 3 Gently load the syringe contents into the loop.
- 4 Leave the syringe in position. The sample will be injected onto the column when the valve is switched to INJECT in the method.

**Note:** If the syringe is taken out before the sample is injected onto the column, self drainage will occur and the loop will be emptied.

### Emptying the sample loop

When emptying the sample loop a buffer volume of approx. 5 times the sample loop volume should be used to flush the loop and ensure that all sample is injected onto the column.

### Storage

**Overnight:** The valve can be left filled with a buffer.

**Weekend or Long time storage:** Flush the valve with water and then fill it with 20% Ethanol.

## 4 Maintenance



**WARNING!** When using hazardous chemicals, take all suitable protective measures, such as wearing protective glasses and gloves resistant to the chemicals used. Follow local regulations and instructions for safe operation and maintenance of the system.

**CAUTION!** Only spare parts approved or supplied by GE Healthcare may be used for maintaining and servicing the valve.

Period	Action
Every 12 months or when required	Change channel plate and distribution plate

### *Cleaning-in-place*

Pump a cleaning or sanitizing agent through the valve. The standard recommendation is to pump 1 M NaOH for 30 minutes and then wash out with buffer.


### *Changing channel plate and distribution plate*



**WARNING!** NaOH is injurious to health. Avoid spillage.

A replacement kit, Valve Kit INV-907, is available, see ordering information in Reference Information.

- 1 Ensure that the valve is in position 1 and then disconnect it from the pump.
- 2 Remove the 4 screws on the front using the supplied 3 mm Allen key. Loosen each one equally in turn so the distribution plate comes off parallel to the valve body.



## 4 Maintenance

- 3 Slide the screws out.
- 4 Remove the distribution plate containing the ports.
- 5 Remove the old channel plate and insert a new one.
- 6 Remount a new distribution plate so that the text 3 is horizontal and to the right of the central tubing connection. Using the Allen key, tighten the 4 screws in turn, a little at a time, until the distribution plate is fixed to the valve body.



This symbol indicates that the waste of electrical and electronic equipment must not be disposed as unsorted municipal waste and must be collected separately. Please contact an authorized representative of the manufacturer for information concerning the decommissioning of equipment.



## 5 Troubleshooting

---

If the suggested actions do not correct the fault, call GE Healthcare.

Fault	Action
<b>The valve is not switching</b>	<ol style="list-style-type: none"> <li>1 Check the connection to the pump. The valve should be connected to the UniNet 2 socket, not the UniNet 1 socket.</li> <li>2 Check the ID-switch on the valve. The ID number should correspond to the number set in UNICORN.</li> <li>3 Check the UniNet cable and replace if required.</li> </ol>
<b>The valve is switching to wrong position</b>	<p>The valve parts may have been incorrectly assembled after replacement.</p> <ol style="list-style-type: none"> <li>1 Check that the distribution plate marking 3 is horizontal.</li> </ol>
<b>External leakage</b>	<ol style="list-style-type: none"> <li>1 Check the tubing connections. Tighten or replace if required.</li> </ol>
<b>Internal leakage</b>	<p>Internal leakage are easy to detect at the small hole on the underside of the valve body.</p> <ol style="list-style-type: none"> <li>1 Internal valve parts may be worn. Change channel plate and distribution plate according to section 4.</li> </ol>
<b>High back pressure (clogging)</b>	<ol style="list-style-type: none"> <li>1 Do cleaning-in-place according to the instructions in section 4.</li> <li>2 Change channel plate and distribution plate according to section 4.</li> </ol>
<b>Other faults</b>	Contact GE Healthcare.



This symbol indicates that the waste of electrical and electronic equipment must not be disposed as unsorted municipal waste and must be collected separately. Please contact an authorized representative of the manufacturer for information concerning the decommissioning of equipment.


## 5 Troubleshooting

## 6 Reference information

### Description

The valve consists of two main parts:

- Housing which encloses the motor and electronics.
- Valve body with a rotating central core.



As the channel plate is turned by the motor, different ports are connected.

Valve switching is controlled from UNICORN by reading the actual position of the channel plate.

The geometry of the valve assures that the flow path is completely swept so that solvent or sample "memory effect" is virtually non-existing. The valve rotates the shortest way to the next position (max. 180°). The material used in the switching parts ensures both long mechanical and chemical lifetime.

The valve housing contains no user replaceable items.

### Technical specifications

#### Operating data

Max Flow rate	100 ml/min
Max Pressure	25 MPa (250 bar, 3600 psi)
Back pressure	<50 kPa at 100 ml/min with water
Leakage	<0.1 µl/min at 25 MPa
pH stability range	1–13, 1–14 (<1 day exposure)
Viscosity	Max. 5 cP
Switch time	<260 ms between two adjacent positions
Operating life time	>50 000 cycles, two adjacent positions
Environment	+4 to +40 °C 20–95% relative humidity 84–106 kPa (840–1060 mbar) atmospheric pressure

## 6 Reference information

### Physical data

---

Internal volume	
Pos 1 LOAD	
Port 1–7	9 µl
Port 2–3	5 µl
Port 4–6	8 µl
Pos 2 INJECT	
Port 1–2	7 µl
Port 3–5	6 µl
Port 6–7	9 µl
Pos 3 WASH	
Port 2–4	8 µl
Port 5–7	9 µl
Flow channel diameter	0.8 mm
Valve principle	Motor controlled valve
Functions	Switching 3 positions controlled from UNICORN
Degree of protection	IP 43
Wetted materials	PEEK (polyetheretherketone)
Chemical resistance	The wetted parts are resistant to organic solvents and salt buffers commonly used in chromatography of biomolecules, except 100% ethylacetate, 100% hexane and 100% tetrahydrofuran (THF)
Power requirement	32 V DC ±10% from the system pump
Power consumption	Up to 9 W
UniNet 2 address	0–9
Inlet and outlet tubing	UNF 10–32 2B "Fingertights" for capillary tubing 1/16" outer diameter
Dimensions,	
H × W × D	135 × 80 × 120 mm
Weight	1.2 kg
EMC Standards	This device meets the requirement of the EMC Directive 89/336/EEC through the following harmonized standards: <ul style="list-style-type: none"><li>• EN 61326 (emission and immunity)</li><li>• EN 55011, GR 2, Class A (emission)</li><li>• This device complies with part 15 of the FCC rules (emission). Operation is subject to the following two conditions:<ol style="list-style-type: none"><li>1 The device may not cause harmful interference.</li><li>2 This device must accept any interference received, including interference that may cause undesired operation.</li></ol></li></ul>

---

**Accessories and spare parts**

<b>Item</b>	<b>Quantity per pack</b>	<b>Code no.</b>
Valve INV-907 including one UniNet cable (fill port, needle and syringe holder are not included)	1	18-1108-40
Injection kit INV-907 including fill port needle and syringe holder	1	18-1110-89
Valve kit INV-907 including channel plate and distribution plate	1	18-1109-05
Sample loops		
100 µl	1	18-1113-98
500 µl	1	18-1113-99
1 ml	1	18-1114-01
2 ml	1	18-1114-02
Superloop, 10 ml	1	18-1113-81
Superloop, 50 ml	1	18-1113-82
Superloop, 150 ml	1	18-1023-85
Cable UniNet, 0.7 m	1	18-1109-74
Mounting bracket	1	18-1109-11
Teflon tubing, i.d. 1/8", o.d. 3/16"	3 m	18-1112-47
Tubing connector for 3/16" o.d. tubing	10	18-1112-49
Ferrule for 3/16" o.d.tubing	10	18-1112-48
Stop plug, 5/16"	5	18-1112-50
Stop plug, 1/16"	5	18-1112-52
Union Luer female/1/16" male	2	18-1112-51
Union 1/16" female/M6 male	6	18-1112-57
Union M6 female/1/16" male	8	18-1112-58
PEEK tubing, i.d. 0.75 mm, o.d. 1/16"	2 m	18-1112-53
Teflon tubing, i.d. 0.75 mm, o.d. 1/16"	2 m	18-1112-54
PEEK tubing, i.d. 1.0 mm, o.d. 1/16"	2 m	18-1115-83
Fingertight connector 1/16"	10	18-1112-55





www.gehealthcare.com

GE Healthcare Bio-Sciences AB  
Björkgatan 30  
751 84 Uppsala  
Sweden

ÄKTA and UNICORN are trademarks of GE Healthcare companies. GE, imagination at work and GE monogram are trademarks of General Electric Company.

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. GE Healthcare reserves the right, subject to any regulatory and contractual approval, if required, to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation. Contact your local GE Healthcare representative for the most current information.

UNICORN Any use of this software is subject to GE Healthcare Standard Software End-User License Agreement for Bio-Sciences Software Products.

© 2005 General Electric Company – All rights reserved.

GE Healthcare Bio-Sciences AB, a General Electric company.

GE Healthcare Bio-Sciences AB  
Björkgatan 30, SE-751 84 Uppsala, Sweden

GE Healthcare Europe GmbH  
Munzinger Strasse 5, D-79111 Freiburg, Germany

GE Healthcare UK Ltd  
Amersham Place, Little Chalfont, Buckinghamshire, HP7 9NA, UK

GE Healthcare Bio-Sciences Corp  
800 Centennial Avenue, P.O. Box 1327, Piscataway, NJ 08855-1327, USA

GE Healthcare Bio-Sciences KK  
Sanken Bldg. 3-25-1, Hyakunincho, Shinjuku-ku, Tokyo 169-0073, Japan

Asia Pacific Tel: +852 2811 8693 Fax: +852 2811 5251 • Australasia Tel: +61 2 2899 0999 Fax: +61 2 9899 7511 • Austria Tel: 01 576 0616 25 Fax: 01 576 0616 27 • Belgium Tel: 0800 73 888 Fax: 03 272 1637 • Canada Tel: 1 800 463 5800 Fax: 1 800 567 1008 • Central, East, South East Europe Tel: +43 1 982 3826 Fax: +43 1 985 8327 • Denmark Tel: 45 16 2400 Fax: 45 16 2424 • Finland & Baltics Tel: +358 (0)9 512 3940 Fax: +358 (0)9 512 39 439 • France Tel: 0169 35 67 00 Fax: 0169 41 9677 • Germany Tel: 0761 4903 490 Fax: 0761 4903 405 • Italy Tel: 02 27322 1 Fax: 02 27302 212 Japan Tel: 81 3 5331 9336 Fax: 81 3 5331 9370 • Latin America Tel: +55 11 3933 7300 Fax: +55 11 3933 7304 • Middle East and Africa Tel: +30 210 9600 687 Fax: +30 210 9600 693 • Netherlands Tel: 0165 580 410 Fax: 0165 580 401 • Norway Tel: 815 65 555 Fax: 815 65 666 • Portugal Tel: 21 417 7035 Fax: 21 417 3184 • Russia & other C.I.S. & N.I.S. Tel: +7 (095) 232 0250,956 1137 Fax: +7 (095) 230 6377 • South East Asia Tel: 60 3 8024 2080 Fax: 60 3 8024 2090 • Spain Tel: 93 594 49 50 Fax: 93 594 49 55 • Sweden Tel: 018 612 1900 Fax: 018 612 1910 Switzerland Tel: 0848 802812 Fax: 0848 802813 • UK Tel: 0800 616 928 Fax: 0800 616 927 • USA Tel: +1 800 526 3593 Fax: +877 295 8102



imagination at work

Instructions 56-8101-01 AE